

	Catatan Hasil Kalibrasi Internal <i>Internal Calibration Record</i> Fluorescence Spectrophotometer	No. : F-PM-01-21
		Rev. : 00
		Berlaku / Date : 3 September 2014

No. Fasilitas / Facility No.	:	Lokasi / Location	:
No. Kode Kalibrasi / Calibration code	:	Bidang / Department	:
Merek / Brand	:	Suhu ruangan / Room temp.	:
Tipe / Type	:	Kelembaban udara / RH	:
No seri / Serial Number	:	No. Protap Cara Kalibrasi / SOP	:
Kapasitas / Capacity	:	Nama Petugas / Name	:

Kalibrator yang digunakan Calibrator	Tanggal Kalibrasi Kalibrator Cal. date of Calibrator	Tgl. Rekalibrasi Kalibrator Recal. date of Calibrator

1. Pengujian Sensitifitas dan Drift / Measurement of Sensitivity and Drift Tanggal/ Date: 10/5/2013

Pengukuran/ Measurement	Hasil/ Result	Syarat/ Requirement	Kesimpulan/ Conclusion	Paraf Petugas/ Signature	Paraf Asman/ Sign. of Asst. Manager
1.		≥ 450			
2.		< 1.5%			

2. Pengujian Linearitas / Measurement of Linearity Tanggal/ Date: 10/5/2013

Penimbangan / Weight	0.02560 gram	
No.	Kepekatan larutan (µg/ml) / Concentrations (µg/ml)	Hasil Pengukuran (Emisi) / Result (Emission value)
1		
2		
3		
4		
5		
6		
Linearitas / Linearity	r = 0.99995	
Syarat/ Requirement	≥ 0.999	
Kesimpulan/ Conclusion		
Paraf Petugas/ Signature		
Paraf Asman/ Sign. of Asst. Manager		