

	<b>Catatan Hasil Kalibrasi Internal</b> <i>Internal Calibration Record</i> <b>Rotavapor</b>	<b>No.</b> : F-PM-01-29
		<b>Rev.</b> : 00
		<b>Berlaku / Date</b> : 3 September 2014

<b>No. Fasilitas /</b> <i>Facility No.</i>	:	<b>Lokasi /</b> <i>Location</i>	:
<b>No. Kode Kalibrasi /</b> <i>Calibration Code</i>	:	<b>Bidang /</b> <i>Department</i>	:
<b>Merek /</b> <i>Brand</i>	:	<b>Suhu Ruangan /</b> <i>Room Temp.</i>	:
<b>Tipe /</b> <i>Type</i>	:	<b>Kelembaban ruangan /</b> <i>Humidity</i>	:
<b>No. seri /</b> <i>Serial number</i>	:	<b>No. Protap Cara Kalibrasi /</b> <i>SOP</i>	:
<b>Kapasitas /</b> <i>Capacity</i>	:	<b>Nama Petugas /</b> <i>Name</i>	:

<b>Kalibrator yang digunakan</b> <i>Calibrator</i>	<b>Tanggal Kalibrasi Kalibrator</b> <i>Cal. date of Calibrator</i>	<b>Tanggal Rekalibrasi Kalibrator</b> <i>Recal. date of Calibrator</i>

**1. Kalibrasi Kecepatan / Calibration of Speed**

<b>Pengujian ke- /</b> <i>Test</i>	<b>25 RPM</b>	<b>50 RPM</b>
<b>1</b>		
<b>2</b>		
<b>3</b>		
<b>4</b>		
<b>Rata-rata / Average</b>		
<b>Syarat / Requirement</b>	<b>24,0 – 26,0 rpm</b>	<b>48,0 – 52,0 rpm</b>
<b>Kesimpulan / Conclusion</b>		
<b>Paraf petugas / Signature</b>		
<b>Paraf Asman / Sign. of Asst. Man</b>		

	<b>Catatan Hasil Kalibrasi Internal</b> <i>Internal Calibration Record</i> <b>Rotavapor</b>	<b>No.</b> : F-PM-01-29
		<b>Rev.</b> : 00
		<b>Berlaku / Date</b> : 3 September 2014

**2. Kalibrasi Termometer Suhu Uap / Calibration Thermometer of Vapor temp.**

Suhu Pengujian / Temperature Test (°C)	Termometer Standar / Reference Thermometer (°C)	Termometer Suhu Uap / Thermometer of Vapor temp.(°C)	Selisih / Diff.. (°C)
30			
35			
40			
45			
50			
55			
60			
65			
70			
<b>Syarat / Requirement</b>	<b>Penyimpangan maksimal 2 °C</b>		
<b>Kesimpulan / Conclusion</b>			
<b>Paraf petugas / Signature</b>			
<b>Paraf Asman / Sign. of Asst. Man</b>			

**3. Kalibrasi Termometer Suhu Waterbath/ Calibration Thermometer of Waterbath temp.**

Suhu Pengujian / Temperature Test (°C)	Termometer Standar / Reference Thermometer (°C)	Termometer Suhu Uap / Thermometer of Vapor temp.(°C)	Selisih / Diff.. (°C)
30			
35			
40			
45			
50			
55			
60			
65			
70			
<b>Syarat / Requirement</b>	<b>Penyimpangan maksimal 2 °C</b>		
<b>Kesimpulan / Conclusion</b>			
<b>Paraf petugas / Signature</b>			
<b>Paraf Asman / Sign. of Asst. Man</b>			