

	Catatan Hasil Kalibrasi Internal <i>Internal calibration Note</i> Piknometer		No. : F-PM-01-33
			Rev. : 00
			Berlaku : 3 September 2014

No. Fasilitas / Facility No.	:	Lokasi / Location	:
No. Kode Kalibrasi / Calibration code	:	Bidang / Department	:
Merek / Brand	:	Suhu ruangan / Room temp.	:
Tipe / Type	:	Kelembaban udara / RH	:
No seri / Serial Number	:	No. Protap Cara Kalibrasi / SOP	:
Kapasitas / Capacity	:	Nama Petugas / Name	:

Kalibrator yang digunakan Calibrator	Tanggal Kalibrasi Kalibrator Cal. date of Calibrator	Tanggal Rekalibrasi Kalibrator Recal. date of Calibrator

Tanggal/ Date	Bobot Pikno/ Weight of Pikno (g)	Bobot Pikno + air/ Weight of Pikno + Water (g)	Bobot Air/ Weight of water (g)	Suhu Air/ Temp. of water (°C)	Factor	Volume (ml)	Kesim- pulan/ Conclusion	Paraf Petugas/ Signature	Paraf Asman/ Sign. of Asst. Manager
A	B	C	D	E	F	G	H	I	J

Catatan / Note :

Keterangan : D = C – B

G = D x F