

	<b>Catatan Hasil Kalibrasi Internal</b> <i>Internal Calibration Note</i> <b>Granulate Flow Tester</b>	No. : F-PM-01-34
		Rev. : 00
		Berlaku/ Date : 3 September 2014

<b>No. Fasilitas /</b> <i>Facility No.</i>	:	<b>Lokasi /</b> <i>Location</i>	:
<b>No. Kode Kalibrasi /</b> <i>Calibration code</i>	:	<b>Bidang /</b> <i>Department</i>	:
<b>Merek /</b> <i>Brand</i>	:	<b>Suhu ruangan /</b> <i>Room temp.</i>	:
<b>Tipe /</b> <i>Type</i>	:	<b>Kelembaban udara /</b> <i>RH</i>	:
<b>No seri /</b> <i>Serial Number</i>	:	<b>No. Protap Cara Kalibrasi /</b> <i>SOP</i>	:
<b>Kapasitas /</b> <i>Capacity</i>	:	<b>Nama Petugas /</b> <i>Name</i>	:

<b>Kalibrator yang digunakan</b> <i>Calibrator</i>	<b>Tanggal Kalibrasi Kalibrator</b> <i>Cal. date of Calibrator</i>	<b>Tanggal Rekalibrasi Kalibrator</b> <i>Recal. date of Calibrator</i>

**1. Kalibrasi Neraca / Calibration of Balance**
**Tanggal/Date :**

<b>No.</b>	<b>Standard Weight (g)</b>	<b>Actual (g)</b>	<b>Syarat / Requirement (g)</b>	<b>Kesimpulan/ Conclusion</b>	<b>Paraf Petugas/ Signature</b>	<b>Paraf Asman/ Sign. of Asst. Manager</b>
1	50		49.95 – 50.05			
2	100		99.90 – 100.10			
3	150		149.85 – 150.15			

**2. Kalibrasi Penunjuk Waktu / Calibration of Timer**
**Tanggal/ Date :**

<b>No.</b>	<b>Nominal (detik/second)</b>	<b>Actual (detik/second)</b>	<b>Syarat (detik/second)</b>	<b>Kesimpulan/ Conclusion</b>	<b>Paraf Petugas/ Signature</b>	<b>Paraf Asman/ Sign. of Asst. Manager</b>
1	10		9 – 11			
2	20		19 – 21			
3	30		29 - 31			

**3. Kalibrasi Sudut Repose / Calibration of Angle of Repose**
**Tanggal/ Date :**

<b>No.</b>	<b>Nominal</b>	<b>Actual</b>	<b>Syarat</b>	<b>Kesimpulan/ Conclusion</b>	<b>Paraf Petugas/ Signature</b>	<b>Paraf Asman/ Sign. of Asst. Manager</b>
1	45°		44° - 46°			
2	45°		44° - 46°			
3	45°		44° - 46°			