

 indofarma	Catatan Hasil Kalibrasi Internal Spectrofotometer HITACHI UH5300	No : F-PM-01-76
		Revisi : 00
		Berlaku : 08 Agustus 2018
		Hlm. : 1/1

No. Fasilitas :	Kapasitas :
Kode Kalibrasi :	Lokasi :
Merek :	Bidang :
Tipe :	No. Protap Kalibrasi :
No. Seri :	Nama Petugas :

Kalibrator yang digunakan	Tgl. Kalibrasi Kalibrator	Tgl. Rekalibrasi Kalibrator

Suhu Ruangan : _____ ° C

Kelembapan Ruangan : _____ % RH

Tanggal Kalibrasi : _____

No.	Parameter	Hasil	Syarat	Kesimpulan	Paraf Petugas	Paraf Spv.
1	<i>Wavelength Accuracy 484,3 nm</i>		± 0,3 nm			
2	<i>Wavelength Accuracy 260.6 nm</i>		± 0,3 nm			
3	<i>Wavelength Accuracy 881.9 nm</i>		± 0,3 nm			
4	<i>Wavelength Repeatability 484,3 nm</i>		± 0,1 nm			
5	<i>Noise Level</i>		<i>Max 0.0001 Abs</i>			
6	<i>Baseline Flatness</i>		± 0,0009 Abs			
7	<i>Baseline Stability</i>		<i>Max 0.0020 Abs</i>			
8	<i>Hardware Check</i>		<i>all Parameter OK</i>			

No. LK : _____

Cibitung,