



Catatan Harian Pemantauan Suhu
Peralatan di Laboratorium Mikrobiologi

No : FPM031
Revisi : 03
Berlaku : 12 Agu 2019
Hal. : 1/1

Nama Alat :	Merek :	Kode Alat :
No. Ruangan :	Penanggung Jawab :	

Bulan & Tahun :					Bulan & Tahun :				
Tgl.	Pagi (°C)	Sore (°C)	Paraf	Ket	Tgl.	Pagi (°C)	Sore (°C)	Paraf	Ket
1					1				
2					2				
3					3				
4					4				
5					5				
6					6				
7					7				
8					8				
9					9				
10					10				
11					11				
12					12				
13					13				
14					14				
15					15				
16					16				
17					17				
18					18				
19					19				
20					20				
21					21				
22					22				
23					23				
24					24				
25					25				
26					26				
27					27				
28					28				
29					29				
30					30				
31					31				

- Catatan: Syarat Suhu Alat :
1. Inkubator Pertumbuhan Jamur (Aqualytic AL 365 & Equitec): 20-25 °C.
 2. Inkubator Pertumbuhan Bakteri (Inkubator 600, Memmert IN450, Aqualytic AL 185 & Kottermann) : 30 – 35 °C.
 3. Inkubator Pertumbuhan Bakteri E.coli (Recitem): 42 – 44 °C.
 4. Cooler : 2 – 8 °C.
 5. Freezer : Minimal –15 °C

Bekasi,

Asman Pengujian Mikrobiologi & IPC